

Nederlands Meetinstituut

Member State
The Netherlands

OIML Certificate N° R76/1992-NL-00.21
Project number 10121312
Page 1 of 2

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1, Dordrecht
Person responsible: W.A.C.M. van Leeuwen

Applicant

Name: Ohaus Corp.
Address: 29 Hanover Road
Florham Park, NJ 07932-0900
U.S.A.

Manufacturer of the certified pattern

Name: Ohaus Corp.
Address: 29 Hanover Road
Florham Park, NJ 07932-0900
U.S.A.

Identification of the certified pattern

Type : Ranger...

$n \leq 6000$ divisions
 $n \leq 3500$ divisions (per weighing range)
 $1.5 \text{ kg} \leq \text{Max} \leq 60 \text{ kg}$
 $e \geq 0.2 \text{ g}$
maximum of two partial weighing ranges

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report, the type-approval certificate and the description with number T5794 and the appertaining documentation folder), with the requirements of the following Recommendation of the International Organisation of Legal Metrology (OIML):

R76
edition 1992
for accuracy class (III)

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval.

Nederlands Meetinsituut
Hugo de Grootplein 1
3314 EG Dordrecht
Telephone +31 78 6332332
Telefax +31 78 6332309

NMI B.V. (Chamber of Commerce Haaglanden
No.27228701)


Subsidiary companies
NMI Certin B.V. (27233418)
NMI Van Swinden Laboratorium B.V. (27228703)
NMI International B.V. (27239176)

This document is issued under the provision that
NMI B.V. nor its subsidiary companies accept
any liability

Reproduction of the complete document is
allowed. Parts of the document may only be
reproduced after written permission



QUALIFIED
BY STERLAB
Reg. nr. L 029



Nederlands Meetinstituut

Member State
The Netherlands

OIML Certificate N° R76/1992-NL-00.21
Project number 10121312
Page 2 of 2

The conformity was established by tests described in the associated test report:
N° R76/1992-NL-00.21, that includes 53 pages.

The issuing authority
W.A.C.M. van Leeuwen
Manager Certification
L.A.

4 October, 2000

The OIML member
G.J. Faber

4 October, 2000

*
**

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.